

Last updated: 21/05/2024 09:31:03

TRO ANALYZER REPLACEMENT CUVETTE

Product group: **661** Product number: **779035**

Accessory: TRO cuvette replacement

Product information

The cuvette in the CLX-XT/XT2 TRO analyzer is a transparent container used to hold the water sample during analysis, ensuring accurate and reliable measurement of total residual oxidants.



Note: The glass cuvette should be inspected.

Check the condition of cuvette and change if it appears badly soiled or discolored.

Check for excessive debris on the inside surface of the glass. It is suggested to keep a spare cuvette to replace when required.

This page is printed from

<https://www.wilhelmsen.com/product-catalogue/products/marine-chemicals/test-kits-and-reagents/water-test-kit---spares-and-consumables/tro-analyzer-replacement-cuvette/>